


# Manual

## SmartGraph for Humlog 10



State: 10.12.2001

Version: V1.0

<b>1</b>	<b>INTRODUCTION TO SMARTGRAPH .....</b>	<b>4</b>
1.1	Manage, Configure.....	4
1.2	The Programme Interface.....	4
1.2.1	Graphs .....	5
1.2.2	Table .....	6
1.2.3	Auto Monitor .....	7
1.2.4	Monitor for Tables and Graphs .....	8
1.3	Process Measuring Instruments .....	9
1.3.1	Read Measuring Instrument.....	9
1.3.2	Recorder .....	9
1.3.3	Delete Measurement Data Memory .....	10
1.3.4	Read Configuration .....	10
1.3.5	Synchronise Clock .....	10
1.4	Analyse and Edit.....	11
1.5	Documentation .....	12
1.6	Organisation of Measurement Data on Hard Disk.....	12
1.7	Upgrade to Extended Functions .....	13
<b>2</b>	<b>MENU COMMANDS .....</b>	<b>14</b>
2.1	File Menu.....	14
2.1.1	Current Unit.....	14
2.1.2	Print .....	14
2.1.3	Printer Set-up.....	14
2.1.4	Page set-up .....	14
2.1.5	Quit .....	14
2.2	Edit Menu.....	15
2.2.1	large time step backwards.....	15
2.2.2	short time step backwards .....	15
2.2.3	Enlarge.....	15
2.2.4	Reduce .....	15
2.2.5	short time step forwards .....	15
2.2.6	large time step forwards.....	15
2.2.7	Data end.....	15
2.2.8	Input date / time range .....	15
2.2.9	Yt Graph or Table .....	16
2.2.9.1	Edit Title.....	16
2.2.9.2	Merge Out/Merge In Column (for tables only).....	16
2.2.9.3	Edit Graph (for graphs only).....	17
2.2.9.4	Value Range Left/Right Y Axis (for graphs only).....	18
2.2.10	Menu Heading Monitor .....	18
2.3	View Menu .....	19
2.3.1	Monitor.....	19
2.3.2	Yt Graph.....	19
2.3.3	Table .....	19
2.3.4	Monitor Menu Headings.....	19
2.4	Measuring Instruments Menu .....	20
2.4.1	Communication.....	20
2.4.2	Manage Measuring Instruments .....	20
2.4.2.1	Edit Unit.....	21

2.4.2.2	Edit Channel .....	23
2.4.3	Remove Measuring Instrument.....	24
2.4.4	Read Measuring Instrument.....	24
2.4.5	Recorder .....	24
2.4.6	Read Configuration .....	24
2.4.7	Synchronise Clock .....	25
2.5	Options Menu.....	25
2.5.1	Protokoll .....	25
3.5.2	Auto Monitor .....	26
2.5.4	Upgrade .....	26
2.5.5	Export Function.....	26
2.5.5.1	Group Channels.....	27
2.6	Help Menu.....	27
2.6.1	Additions to the manual .....	27
2.6.2	Contents.....	28
2.6.3	Using Help .....	28
2.6.4	Info.....	28
<b>3</b>	<b>APPENDIX.....</b>	<b>29</b>
3.1	Keyboard Operation .....	29

# 1 Introduction to SmartGraph

## 1.1 *Manage, Configure*

SmartGraph enables the flexible management of several units by means of identifiers. Measuring instruments can be added at any time. Every measuring instrument can be provided with a description.

The measurement data is automatically archived in Calendar Week files, whereby a sub directory DIXXX is used for each measuring instrument.

For later access to the stored measurement data it is sufficient to select the measuring instrument Current Unit using the name in the File Menu.

For security against hard disk errors, completed Calendar Week Files should be regularly saved to diskette or tape and kept in a safe place.

See also: [2.4.2 Manage Measuring Instruments](#)

## 1.2 *The Programme Interface*

The programme interface provides the most comprehensive presentation possible of all the information, in particular the measurement data.

Measurement data can be viewed via either Graphs or Tables. The programme automatically compiles graphs and tables for the desired measuring instrument on the interface. Information regarding selection of the measuring instrument to be presented and the formulation of the interface can be found under Auto Monitor.

In order to be able to present a greater number of measurement channels the interface is sub-divided where necessary into pages lying on top of one another. These pages are identified as Monitors. Switching between different monitors takes place by selecting from a list on the lower frame. The list disappears if only one monitor is in use.

Different information is presented on the monitors. Correspondingly, there are different monitor types:

- Protocol Monitor: The protocol entries for the preceding time are listed here.
- Monitor for Tables and Graphs

## 1.2.1 Graphs

Graphs enable the presentation of measurement values in trace form on a date/time axis. Up to 10 traces with measurement values from different sources, i.e. the selected channels of various units, can be displayed.

To present measurement values from different units, 2 Y-axes are used (left and right). In this way the measurement values from two different instruments can be presented on a single graph. Dots in the colours of the traces alongside the unit of the relevant Y-Axis facilitate the allocation of the trace to the corresponding Y-Axis.

### Navigation within the Date/Time Range

Corresponding control buttons in the tools panel serve to navigate in the Date/Time range (X-axis). It is possible to move short or long distances in time forwards or backwards, or to enlarge (Zoom) or reduce (Un-zoom) the screen section along the X-axis. Zoom and Un-zoom can also be obtained by double-clicking with the mouse in the area of the graphs. In this case the mouse pointer appears as a magnifying glass. Double-clicking with the left mouse button stretches the time range to the mouse position, double-clicking with the right mouse button switches from the current time period to the next largest time range.

### Context Menu

For quick and easy changes to Graph contents and settings, a context menu is available. This pops up when the right mouse button is activated on the graph interface. The same menu headings are available under the Edit and View menus (under heading Yt-Graph), when the graph is currently selected (bordered in blue).

The following settings can be specified here:

Edit Title, see [2.2.9.1 Edit Title](#)

Edit Graph, see [2.2.9.3 Edit Graph \(for graphs only\)](#)

Value Range Left/Right Y-Axis, see [2.2.9.4 Value Range Left/Right Y Axis \(for graphs only\)](#)

Grid Network Lines to Left/Right Y-Axis

Caption:

If the caption for the currently displayed graphs is not required this can be merged out in order to provide more space for the graphic.

Print

Print Graph on one page. Can also be obtained via menu heading File-Print-Yt-Graph.

## 1.2.2 Table



The Table presents the measurement values of one day. The time is presented in the right column. The measurement value columns can originate from different units. Measurement values that exceed the alarm limits are highlighted in bold and in red text.

For space reasons the short description is always used in the unit and channel lines. The long description is shown in an advice window when the mouse moves over the appropriate cell.

### Navigation within the Time Range:

The data from the current day can be obtained via the vertical screen scroll bar. The same function can be obtained via the Cursor Keys (Arrow up, Arrow down, Page up, Page down) (See also Keyboard Operation).

### Navigation within the Date Range:

Using the control buttons  and  the user can scroll to the previous or next page, this can also be carried out via the key combination Ctrl Arrow Up or Ctrl Arrow Down.

The time ranges of the tables and the graphs are synchronised. The day presented in the Table is always contained in the time range of the Graphs.

### Context Menu:

For quick and easy changing of the contents and settings of tables a Context menu is available, which is activated by pressing the right mouse button. The same menu headings are also available via the sub menus Tables in the Edit Menu and View Menu, if the table has been selected.

The following operations are possible here:

Edit Title


Merge Out Column

Merge In Column

Horizontal Grid Network Lines

Vertical Grid Network Lines

Print Table:

The Table is printed with the data of the current day. This menu heading can also be obtained via the control button  in the tools panel and via the menu heading File Print \_ Print table.

See also:

[3.1 Keyboard Operation](#)

[1.2.4 Monitor for Tables and Graphs](#)

### 1.2.3 Auto Monitor

The Auto Monitor function enables the automatic build up of one or more monitors with graphs and table, for the presentation of all channels of one data source or one unit.

#### Selecting the current unit:

The selection of the current unit that is being presented by the Auto Monitor can take place in two ways

- from the selection list in the tools panel
- from the sub menu under the menu heading File - current unit

The name and identifier of the current unit are displayed in the header of the main window.

#### Setting up the Auto Monitor:

The automatic building and arrangement of the monitors takes place in accordance with the following settings. The settings can be amended using the dialog "Auto Monitor Settings", which is available from the menu heading Options - Auto Monitor.

- De-activate or Switch On Auto Monitor: With the option De-activate no automatic monitors are provided. This option is only available in the Professional version of the software.
- Maximum number of traces per graph: With this the number of traces that are presented in a single graph can be limited. Values between 1 and 10 are permissible. If more channels than the maximum available are present in a data source than can fit into a graph in accordance with your unit, an additional graph is automatically produced.
- Maximum number of graphs per monitor: The number and units of the channels of a data source determine how many graphs are necessary for presentation. If more graphs are required than the maximum number allowed per monitor, additional monitors are automatically produced.
- Use both axes in graphs: When this option is activated, each graph is provided with a left and right Y-axis if necessary, and in this way channels with two different units can be presented. If the option is deactivated, only the left Y-axis is used.
- Table on separate monitor: When this option is activated, the table appears on a monitor without graphs. Otherwise graphs and table are arranged together on one monitor.

After the settings have been changed using the dialog, the automatic construction is carried out again using the current parameters.

If the current unit is changed, the settings that were active at the time of the last presentation of the unit are used to build the monitor.

On monitors that were produced in this way it is not possible to manually add or remove graphs or tables. They can only be merged out and the sequence changed. No traces or columns can be manually added or removed in the graphs and tables. In tables however the Columns can be Merged Out and the sequence changed. The headers of the monitors are automatically compiled from the unit name and a serial number and likewise cannot be amended.

## 1.2.4 Monitor for Tables and Graphs



The monitor is divided into two halves. On the left side up to 9 graphs are presented, on the right side a table. The border between the two sides can be shifted freely using the mouse. The graphs are arranged automatically on the left side in such a way that the space is filled as completely as possible. The arrangement can take place by choice either alongside or on top of one another (see Monitor Sub menu in the View menu and Context menu).

### Selection:

Whenever a graph or table is selected, i.e. various menu headings or control buttons, such as for example Print, these all relate to the selected object. The selected item is highlighted by means of a blue border. An item can be selected by mouse-click or by using the tab key.

### Full Screen/Normal Screen:

Each item (graph or table) can be expanded to full screen size. This then fills the entire monitor. Switching to full screen and back to normal screen size takes place in 4 ways:


- Activation of the control button  or  on the upper right in the object.
- Key combination <Ctrl> + and <Ctrl> -
- Via the Context menu of the monitor
- Via the Monitor Sub menu in the Edit menu.

The menu headings always relate to the selected object.

### Merge In/Merge Out:

Each object (table or graph) can be merged out, i.e. it disappears from the screen, but is not lost and can be merged back in again. The remaining objects are then arranged automatically.

Merge out takes place via :

- Control button  upper right on the object.
- Context menu of the monitor (see below).
- Via the Monitor Sub menu in the Edit menu.

The menu headings always relate to the selected object.

Merge in takes place via the Context menu or the Monitor Sub menu in the Edit menu. The object to be merged in can be selected via the dialog box that pops up, and the position can also be selected where the item is to be introduced.

Note: The mechanism can be used to change the Sequence of the Objects on a monitor, where objects are first merged out and then reintroduced in the desired place.

### Context Menu for Monitor:

All operations that concern the monitor can be activated via the Context menu. This appears when the right mouse button is activated and the mouse is positioned over the index panel of the Monitor or over a free surface of the monitor. The menu headings also appear in the Monitor Sub menu of the Edit menu (Full Screen/Normal Screen, Merge In / Out) or in the Monitor Sub menu of the View menu (Graphs alongside/on top)

## 1.3 Process Measuring Instruments

Access to the Measuring Instruments.

### 1.3.1 Read Measuring Instrument

The memory content of the measuring instrument that is connected to the PC via the serial interface is read and recorded.

Requirement: The correct interface must be set up.  
For set-up use the command Communication in the Measuring Instrument Menu.

Messages:  
Whilst reading is taking place the message "Reading Measuring Instrument" is shown with an indication of the progress of the transmission (percent reading).

If the user has inadvertently selected the interface to which the mouse is connected, SmartGraph produces the message "Interface already open".

If the message "Measuring Instrument not responding" appears, check whether the correct interface has been set up and if measuring instrument and PC are connected using the correct cable.

### 1.3.2 Recorder

With the Recorder function current measurement data from the connected unit is collected, presented and saved in cyclical manner. In contrast to the command 'Read Measuring Instrument' it is not the measurement value memory of the unit that is accessed here, but the most recently captured measurement value directly.

During the recording process the connection to the measuring instrument is constant. The settings for connection type and interface that were defined under Communication are used as the basis for the connection to the measuring instrument. For the Recorder function, only the connection type PC Direct is permissible.

The Recorder is started or stopped via the menu heading 'Measuring Instrument - Recorder - Start/Stop Recorder' or via the corresponding control buttons in the tools panel. Whilst recording is taking place a symbol appears in the Status panel.

During the recording operation the following functions are not available:

- Communication
- Read Configuration
- Read Measuring Instrument
- Delete Measuring Instrument Memory
- Synchronise Clock

The measuring instrument clock is automatically synchronised during the recording process.

The following actions lead to recording being broken off:

- Changing a unit or channel setting with "Manage Measuring Instruments"
- "Remove Measuring Instrument"

The time interval for online recording with the Recorder is taken from the memory rate of the unit in question, this can be set with the command "Manage Measuring Instruments".

The currently measured data can be presented automatically using the function "Automatic Screen with Data Recording"

This function can be switched on and off from the menu heading in sub menu "Measuring Instrument - Recorder" and from the control field in the tools menu.

This enables that during data recording:

- The unit whose data is currently being recorded is presented on the Auto-Monitor.
- The most recently measured data remain within the presentation area of the Graphs or of the Table.

If necessary the screen will be scrolled to the right or down.

When the function is activated, neither the time range in the traces, the presented day in the table nor the current unit for the Auto Monitor can be changed manually.

### **1.3.3 Delete Measurement Data Memory**

After activation by the user the memory for the measurement data is deleted in the connected unit.

The settings for connection type and interface that were defined under Communication are used as the basis for the connection to the measuring instrument.

If no corresponding unit definition has been made then this is done automatically.

### **1.3.4 Read Configuration**

The current configuration of the connected unit is read. The settings for connection type and interface that were defined under Communication are used as the basis for the connection to the measuring instrument.

If no corresponding unit definition has yet been set then this is produced automatically. At the request of the user the current unit (see Auto Monitor) can now be switched over to the connected unit.

### **1.3.5 Synchronise Clock**

The internal clock of the connected unit is synchronised. The settings for connection type and interface that were defined under Communication are used as the basis for the connection to the measuring instrument.


## 1.4 Analyse and Edit

To analyse the measurement data the presentation range can be selected in steps of 10 seconds to one year.

To navigate within the time range of the graphics use the Navigator on the upper left under the main menu. The same functions can be obtained under the menu heading Edit. Zoom and Un-zoom can also be carried out using mouse click. See also: [3.1 Keyboard Operation](#)

Enlarge, Reduce:

As soon as you move the mouse pointer within the display range of a graphics window it changes into a magnifying glass.

To  Enlarge hold the magnifying glass over the required position and click twice with the left mouse button (double click).


The graphic now switches to the next largest size, with the clicked position lying in the centre. (Zoom).

To  Reduce simply click twice with the right mouse button (Un-zoom).


The **Presentation Steps** are:

**Year-Month-Week-Day-6Hours-1Hour-10Minutes-1Minute-10Seconds**


**Roll:**

 small time step backwards


To move the graphic section back to the next smallest Time Step

 mall time step forwards

To move the graphic section forward to the next largest time step

 large time step backwards

To move the graphic section back by an entire Time Step.

 large time step forwards

To move the graphic section forward by an entire time step.

 Data end

To move the graphic section so that the last data set lies in the presentation range.

 Input date / time range

A Dialog Box appears, in which the time range for the presentation can be input. The difference between start time and stop time must be at least 10 seconds and may be a maximum of one year.

Each operation in a given graphic window will also be carried out in all other graphic windows. The graphic windows therefore always display the same time section. (When superimposed on top of one another the results can also be compared very effectively.)

## **1.5 Documentation**

Printing of tables and graphs:

Printout on the installed Windows printer with full printer resolution.  
(See [2.1.2 Print](#) und [2.1.3 Printer Set-up](#) in the File Menu).

## **1.6 Organisation of Measurement Data on Hard Disk**

SmartGraph provides its own sub directory for each measuring instrument **dlxxx** (xxx =identification) in the SmartGraph programme directory. Sub directories (dlxxx) are produced for the channels of these measuring instruments under the associated instrument directories.

The measurement data is saved in Calendar Week files "**kwjjnn.sgr**". Here **jj** signifies the year and **nn** the number of the calendar week.

When a measuring instrument is deleted the subdirectory DLxxx of the corresponding instrument (xxx =identification) is renamed DLxxxBAK. The files are not deleted. If a measuring instrument has been inadvertently deleted, the procedure can be reversed in the following way: Quit SmartGraph and copy **All Files** from the directory **DLxxxBAK** into the directory **DLxxx** (e.g. using the File Manager). The next time SmartGraph is started all the files will then be available.

## **1.7 Upgrade to Extended Functions**

### Functions of the Professional Version

- Instead of the standard measurement units (e.g. temperature in °C and humidity in %RH) the individual channels of the measuring instrument can be converted to other measurement units (e.g. temperature in °F or dew point temperature, etc.)
- Export of measurement data to text files.

### Release via Key-Code

By acquiring an individual Key-Code the above functions can be easily released.

To do this, select the menu heading Upgrade/Order in the Options Menu. Fill out the fields in the dialog box "Quotation/Order for Release Code". Select Print to print out an Order form. Fax the order to the given Fax number and you will shortly receive the Key-Code for the desired release of the appropriate version.

If you are in agreement with the Licence conditions select the menu heading Upgrade/Release in the Options Menu. Enter the Key-Code. The additional functions can now be used.

Approval release for the additional functions must be given in accordance with the number of installed units. If more units are present the additional functions are deactivated.

## 2 Menu Commands

### 2.1 File Menu

The File menu contains the following commands:

#### 2.1.1 Current Unit

Selection of the unit to be presented on the Auto Monitor

#### 2.1.2 Print

Print graphic or table

You may choose whether to print the selected object (graph or table) or a Snapshot (complete visible monitor).

Tip: When printing, the logo saved in the file LOGO.BMP in the sub directory \Support is printed in the right upper corner. If you replace this file you can use your own logo. This can be produced using the programme PaintBrush, for example.

#### 2.1.3 Printer Set-up

Set up options on the installed Windows printer.

Select options for your printer here.

With most printers you have the choice between various graphic resolutions. It is true that a higher resolution increases quality, however the time required for printing is longer. Try various resolutions to find the best solution.

Tip: For printing graphs the format "Landscape" is best suited, for printing tables the format "Portrait" may be better suited, depending on the number of columns.

#### 2.1.4 Page set-up

Set up title and page borders

The dialog box 'Page Set-up' appears with the following input fields:

Title, Sub Title

Fixed page header, that will be printed on every page (e.g. for user data).

Page Borders

Input the desired border in cm here. (The page size must be selected using the command Printer Set-up.)

Tip: A logo appears in the top right corner of every printed page. This logo is read from the file Logo.bmp, which is located in the subdirectory \Support of SmartGraph Index. This file can be replaced if a different logo is required. If no logo is required the file can be renamed or deleted.

#### 2.1.5 Quit

Quit SmartGraph.

## 2.2 Edit Menu

The Edit menu contains the following commands:



### 2.2.1 large time step backwards

To move the graphic section back by an entire Time Step.



### 2.2.2 short time step backwards

To move the graphic section back by the next smallest Time Step



### 2.2.3 Enlarge

The content of the graphic window is enlarged (Zoom)

Enlarge the graphics window to the next time stage, whereby the centre is enlarged.

A double-click with the left mouse button whilst the mouse pointer is within the area of a graphic window has the same function (mouse pointer =magnifying glass), the only difference is that here the centre of the enlargement can be defined by the position of the magnifying glass.

Stages are: Year-Month-Week-Day-6Hours-1Hour-10Minutes-1Minute-10Seconds.



### 2.2.4 Reduce

The content of the graphic window is reduced to the next smallest size (Un-zoom)

Reduce the graphics window to the next time stage.

The identical function is obtained by double-clicking with the right mouse button, where the mouse pointer becomes a magnifying glass.

Stages are: Year-Month-Week-Day-6Hours-1Hour-10Minutes-1Minute-10Seconds.



### 2.2.5 short time step forwards

To move the graphic section forward by the next smallest time step



### 2.2.6 large time step forwards

To move the graphic section forward by an entire time step



### 2.2.7 Data end

To move the graphic section so that the last data set lies in the presentation area.



### 2.2.8 Input date / time range

A Dialog Box appears, in which the time range for the presentation can be input.

The difference between start time and stop time must be at least 10 seconds and may be a maximum of one year.

The other menu headings are relevant to the selected object:

When a monitor is in use, to which these operations are not applicable (e.g. Protocol monitor), these

menu headings do not appear.

## **2.2.9 Yt Graph or Table**

These menu headings are also available in the context menu, this pops up when the right mouse button is activated over the surface of a graph.

The following settings can be specified here:

### **2.2.9.1 Edit Title**

Whenever an object is produced (Graph, Table, etc.) a serial number is always provided. The Standard Title is compiled from the object description and this number.

With the menu heading Edit Title a dialog box appears in which the user can enter a description of his choice for this object instead of the Standard Title. The title becomes effective when the field Special Title is marked and the dialog is ended with OK. An empty title is not permitted.

The user can revert to the Standard Title at any time by marking this field and ending the dialog with OK.

### **2.2.9.2 Merge Out/Merge In Column (for tables only)**

#### Merge Out Column:

A dialog box appears in which the desired column can be selected. By activating with 'OK' the column is removed from the table, but not deleted.

#### Merge In Column:

This menu heading is only available if columns have earlier been merged out. A dialog box appears in which the measurement unit to be inserted can be selected. With the second selection box the column can be selected before which the additional column is to be inserted, or the column can be added at the end.

Note: The mechanism can be used to change the sequence of the columns in a table, whereby columns are first merged out and then inserted in the desired place.

### 2.2.9.3 Edit Graph (for graphs only)

The dialog enables the settings of traces within graphs to be changed.

The available traces are shown in a list on the left side of the dialog. The entry in the list that belongs to the current trace is highlighted, i.e. selected. If another trace is to be edited this can be selected by clicking on the corresponding entry.

The parameters that can be set on the selected trace are shown on the right side of the dialog:

- Trace visible: If this setting is activated the trace appears in the graph and in the caption, otherwise the presentation is suppressed. In so doing the trace is not deleted from the graph but can be made visible again with all the previous settings.

- Show alarm limits: If this setting is activated the alarm limits of the selected trace appear on the graph in the colour of this trace. This button is only available when the 'Trace Visible' setting is activated.

- Colour of the trace: By clicking with the left mouse button a list appears with the 16 standard colours. Colour selection takes place by clicking on the corresponding entry. If a colour different to the standard colours is desired, clicking with the right mouse button on the selection element starts a colour selection dialog. The trace and the corresponding caption entry are then shown in the selected colour.

#### Control Buttons:

- Transfer: With "Transfer" all changes specified since the start of the dialog, or since the last "Transfer", are saved. Thereafter, the changes cannot be reversed by activating "Cancel".

- Cancel: The dialog is ended without saving the changes that were specified since the beginning of the dialog or since the last activation of "Transfer". These changes are thereby lost.

- Close: The changes are saved in the same way as for "Transfer". The dialog is then ended.

#### **2.2.9.4 Value Range Left/Right Y Axis (for graphs only)**

The dialog enables the value range of the Y-axis to be changed within the graphs.

The minimum or maximum of the range can be defined automatically by the programme or manually by the user.

##### Automatic Setting:

The automatic setting is selected by activating the relevant upper option field. The corresponding value is shown alongside on the right. The maximum value of the automatic setting is the highest of all the values of the presentation ranges of all presented channels (traces). The minimum value corresponds to the lowest of all the values of the presentation ranges of all the presented channels.

##### Manual Setting:

The manual setting is effective when the relevant lower option field is activated and the desired value is registered in the input field alongside.

##### Control Buttons:

- OK: With "OK" all changes are saved. The dialog is closed.
- Cancel: The dialog is closed without saving the changes.

#### **2.2.10 Menu Heading Monitor**

##### **Menu Heading Monitor**

This sub menu appears when a Monitor for Graphs and Tables is currently visible.

##### Object Full Screen

The selected object is enlarged to full screen size

##### Object Normal Screen

The enlarged object is returned to its normal size

##### Merge In / Merge Out object

The selected object is merged out, but not deleted. With 'Merge in Object' a dialog box appears with all the previously merged out objects. An object can be selected here that is to be merged back in again.

When a monitor is in use, to which these operations are not applicable (e.g. Protocol monitor), these menu headings do not appear.

## 2.3 View Menu

The menu heading View contains the appropriate sub menu headings depending on the object selected:

### 2.3.1 Monitor

This sub menu appears when a Monitor for Graphs and Tables is actually visible.

To select whether graphs are to be arranged alongside one another or on top of one another

- Arrange graphs alongside one another
- Arrange graphs on top of one another

### 2.3.2 Yt Graph

Grid Network Lines to the Left Y-Axis:

This menu heading is activated and de-activated by clicking on it. When the menu heading is activated the grid network lines for scaling the left Y-axis are shown on the graph.

Grid Network Lines to the Right Y-Axis:

This menu heading is activated and de-activated by clicking on it. When the menu heading is activated the grid network lines for scaling the right Y-axis are shown on the graph.

Caption:

If the caption for the currently displayed graphs is not required this can be merged out in order to provide more space for the graphic.

These menu headings are also available in the context menu, which pops up when the right mouse button is activated over the surface of a graph.

This sub menu appears when a Graph is actually selected.

### 2.3.3 Table

This sub menu is available if Table is currently selected.

Horizontal Grid Network Lines: Merge In / Out

Vertical Grid Network Lines: Merge In / Out

### 2.3.4 Monitor Menu Headings

With these one can switch between the monitors. The monitor presented is suitably checked ✓ .

## 2.4 Measuring Instruments Menu

### 2.4.1 Communication

A Dialog Box appears for selection of the interface.

Mark the interface to which you wish to connect the measuring instrument and close the dialog box with 'OK'.

**Note:**

If you have wrongly selected the interface to which your mouse is connected, SmartGraph produces the message '**Interface already open**' on every attempt to open the interface.

### 2.4.2 Manage Measuring Instruments

This command opens the dialog box Manage Measuring Instruments

The dialog enables the settings of units, sub modules and channels to be changed.

The available elements that can be modified are presented in a tree diagram in the left part of the dialog. The tree diagram functions in a similar way to that in Windows Explorer.

The desired element can be selected here by Clicking on the corresponding entry.

By Double-clicking on an entry the associated list of the subordinate entries can be expanded or merged out.

In this way all elements of the instrument structure are directly selectable and yet complex structures can also still be presented comprehensively.

The selected entry appears highlighted in colour.

On the right side of the dialog the parameters of the selected element can be worked on.

Information concerning this can be found under

- Edit Unit
- Edit Channel

Control Buttons:

- OK: With "OK" all changes that were made since the beginning of the dialog are saved. The dialog is ended.
- Cancel: The dialog is ended without the changes being saved. All inputs that were made since the beginning of the dialog are lost.

### 2.4.2.1 Edit Unit

The following parameters of a unit can be changed:

#### Identifier:

This number identifies a unit uniquely within the SmartGraph software. It defines the lowest level of the database index in which the unit data is recorded. An identification of 5, for example, belongs to a subdirectory \DL005.

The identifier can only be changed by adding a new unit within SmartGraph. Later, in particular once data has already been recorded for this unit, no further change is possible.

#### Description:

The description of a unit can be freely selected. Within the software the unit is interrogated using this name and measurement traces, columns in tables etc., are allocated with this name.

#### Short Description:

The short description can also be freely selected and serves the same purpose as the description. It is used in places where a long name would for space reasons be unsuitable on the visualisation interface, e.g. on graph captions or column headings in tables.

The short description should therefore consist of only a few characters. If necessary, abbreviations should be used.

#### 2.4.2.1.1 Humlog10 Unit Parameters

The setting of parameters takes place on the following indices:

##### - Information:

The descriptions can be entered here and internal unit information such as revision status can be read.

##### - Logging Mode:

The parameters for data capture or recording are set here

Memory organisation: In Start/Stop mode data recording is ended when the memory is full. In ring mode when the memory is full the oldest data is overwritten

Memory mode: At each memory time the following values can be saved:

Average value of the sampling values across the memory interval

Maximum of the sampling values across the memory interval

Minimum of the sampling values across the memory interval

At least one value must be selected.

Sampling and memory interval: The memory interval must always be greater than or the same as the sampling interval.

##### - Pre-selection Operation:

When pre-selection operation is activated the data memory is interrupted and only continues at the pre-set time. Times in the past are ineffective.

##### - Display

Either the current status of the unit clock or the memory start time (when pre-selection is activated) is presented on the unit display. Date and time are shown alternately. The duration of each indication is input here. Values between 1 and 15 seconds are permitted.

##### - Sensor Units:

The sensors attached to the two channels are displayed here.

Settings: Using the control buttons the set-up mask for the corresponding channel can be accessed (see Humlog10 Channel Parameters).

Change Sensor Units: With this control button the measurement units that are captured in the individual unit channels can be changed. The dialog box Change Sensor Units appears. This setting possibility is only available when the "Professional" programme version has been supplied for the available number of units (see Upgrade).

Detailed instructions on the setting possibilities are contained in the unit manual.

Notes:

- If parameters that are required within the unit itself have been changed, by activating 'OK' connection is automatically made to the unit in order to transfer the new values. For this purpose the unit must be connected to the PC and the corresponding serial interface must be set up under "Measuring Instrument - Communication
- If changes have been made to a unit other than through SmartGraph (e.g. new software version or different sensor hardware) this change of condition must first be made known to SmartGraph using the command 'Measuring Instrument \_ Read Configuration' , before parameters can be changed.

### 2.4.2.2 Edit Channel

The following parameters of a channel can be changed:

Description:

The description of a channel can be freely selected. Up to 79 characters are permissible. Within the software the channel is interrogated using this name, and measurement traces, columns in tables etc., are allocated with this name.

Short Description:

The short description can also be freely selected and serves the same purpose as the description. In this case a maximum of 20 characters is permissible. It is used in places where a long name would for space reasons be unsuitable on the visualisation interface, e.g. on graph captions or column headings in tables.

The short description should therefore consist of only a few characters. If necessary, abbreviations should be used.

Unit:

This describes the physical unit of the measurement channel or the channel being computed. Up to 20 characters are possible. The unit is defined internally within the software and cannot be changed by the user.

Measurement Range:

The maximum and minimum of the value range of the measurement channel or calculated channel are presented here. The values are defined internally within the software and cannot be changed.

Resolution, Decimal Place:

The decimal places are specified here, i.e. the number of places after the decimal separator that will be used when presenting values in tabular form. Values between 1 and 10 are permitted.

Note: The value can be selected independently of the measurement accuracy or resolution of the connected unit, however it should relate sensibly to the unit's specification.

Alarm Limits:

The upper and lower limit values for the alarm presentation can be selected here. These limits, and the exceeding thereof, are presented in the graph or in tables, if the corresponding option has been set up.

Presentation Range in Graphs:

It can be specified that in graphs the presentation range (Y-range) be automatically defined. The given maximum and minimum are the basis for the automatic definition. See also the menu heading Value Range Left/Right Y-Axis in the context menu of graphs.

#### 2.4.2.2.1 Humlog10 Channel Parameters

Setting of the parameters takes place on the indices:

- Information:

The descriptions and an offset for drift correction can be input here.

- Alarm:

Input the parameters for alarm activation.

On the alarm limits the maximum must be greater than or the same as the minimum.

Alarm limits outside the measuring range of the sensor will be limited to the measuring range.

- Display

see Edit Channel

Detailed instructions on the setting possibilities are contained in the unit manual.

Notes:

- If parameters that are required within the unit itself have been changed, by activating 'OK' connection is automatically made to the unit in order to transfer the new values. For this purpose the unit must be connected to the PC and the corresponding serial interface must be set up under Measuring Instrument Communication
- If changes have been made to a unit other than through SmartGraph (e.g. new software version or different sensor hardware) this change of condition must first be made known to SmartGraph using the command Measuring Instrument Read Configuration, before parameters can be changed.
- 

#### 2.4.3 Remove Measuring Instrument

With this command a measuring instrument can be deleted from management by SmartGraph. After removal the saved data and configuration information are no longer available.

When called up the dialog box 'Remove Measuring Instrument' appears.

Select the desired measuring instrument from the list. On activating 'OK', the selected measuring instrument is removed, on 'Cancel' no operation is carried out.

Instructions regarding reproduction of the data can be found under Organisation of Measurement Data on Hard Disk

#### 2.4.4 Read Measuring Instrument

To read measuring instrument memory and save data.

See also: [1.3.1 Read Measuring Instrument](#)

#### 2.4.5 Recorder

With the Recorder function current measurement data from the connected unit are collected, presented and saved in a cyclical manner.

See also: [1.3.2 Recorder](#)

#### 2.4.6 Read Configuration

To read the current configuration of the connected unit.

See also: [1.3.4 Read Configuration](#)

### **2.4.7 Synchronise Clock**

The internal clock of the connected unit is synchronised with the PC clock.

See also: [1.3.5 Synchronise Clock](#)

## **2.5 Options Menu**

Commands:

### **2.5.1 Protokoll**

Selection of the protocol options in a dialog box.

This dialog box contains protocol options that can be selected:

#### Short Protocol

Not all messages are written into the protocol.

#### Automatic Management of Protocol Files

The size of the protocol file is limited. If the size is exceeded the old file is renamed.

The new file name contains the date of the last entry.

#### Maximum Length of Protocol File

The maximum length of the protocol file can be selected between 10 Kbytes and 9999 Kbytes. This setting possibility is only available when the option Automatic Management of the Protocol Files has been selected.

#### Protocol only in Auto Mode

Messages are only written to the protocol file in Auto Modus.

When a protocol file is called up, an additional monitor appears automatically, on which the messages are produced.

Auto Mode is only active when the recorder is running, or when an automatic interrogation is included.

### 3.5.2 Auto Monitor

Dialog for editing the Auto Monitor characteristics.

See also: [1.2.3 Auto Monitor](#)

### 2.5.3 Language



To change the current language. This function can be called up under the language button


### 2.5.4 Upgrade

Sub menu to order and input a release code.

See also: [1.7 Upgrade to Extended Functions](#)

### 2.5.5 Export Function

With the Export function, measurement data recorded in the SmartGraph database can be made available in the form of tables in text files. These text files can be further processed in other applications (e.g. word processing or spreadsheet calculation).

The function is activated from the menu heading 'Options \_ Export' or the corresponding control button  in the tools panel. For Humlog10 the function is only available when the Professional Version (Upgrade) has been released.

On activation a dialog appears, by means of which all the parameters for export are defined in a series of several sequential steps. Steps that have already been carried out can be reprocessed at any time using the control button 'back'.

#### Step 1: Input Export Time Range

The date and time for start and finish of the exported time range can be input. For this the finish must be at least 1 second after the start. Activation of the option field 'export without date' has the effect that the date column is omitted. This option is used when, for example, the data of just one day is exported.

#### Step 2: Selection of the Channels to be exported

The channels are shown in the list with the name of the unit and the channel. In the table one column corresponds to each channel.

Selection can be made in two ways:

- Activation of the control button 'Channels of Current Unit' has the effect that Auto Monitor includes all the channels of the 'Current Unit'.
- On activation of 'Group Channels' the dialog [Group Channels](#) appears, by means of which single or multiple channels from various different units can be selected.

#### Step 3: Channel information to be exported

In addition to the measurement data, information can be taken for further processing via the exported channels. When exported into the text file this information is arranged in tabular form over the measurement data. A line is available for each item of information. The table containing the measurement data is separated by the word 'Data'.

The following information is available and can be selected by activating the option field in the list.

- Channel number: serial number for the channels in the corresponding column
- Data source: full name of the associated instrument
- Data source short: short name of the associated instrument
- Channel: full name of the channel
- Channel short: short name of the channel
- Sensor type: Identification code for the sensor type (not relevant for Humlog10)
- Unit: Physical measurement unit
- Min: Minimum of the measurement range
- Max: Maximum of the measurement range
- Alarm min: lower alarm limit
- Alarm max: upper alarm limit

#### Step 4: File name for the text file

The file name including its path can be input directly or via a File Selection Dialog, which can be called up via the control button 'Search'. If an existing file is selected, the option 'Add to existing file' has the effect that the measurement data are added. In this case, channel information (Step 3) is not exported. If the option is not activated if necessary an existing file will be overwritten. If an unavailable file name is selected then this option field does not apply.

#### Step 5: Format settings for text output.

The following settings can be specified:

- Date format: Date separator, Sequence, Year in 2 or 4 characters
- Time: Separator
- Decimal setting ( . or , ) for measurement values
- Separator between columns (e.g. channels). May not be the same as the separator for date and time.
- Character for line end (e.g. separation of data sets)
- Character string to be entered for erroneous or invalid measurement values  
e.g. "---" or "MISSING". The character string must not contain the column separator. An empty character string is permitted.

The export is executed using the control button 'Ready'.



### **2.5.5.1 Group Channels**



The dialog enables the selection and grouping of several channels from any data source. The selected channels are then presented, for example, in a table.

The available data sources (units) are presented in a list on the left hand side of the dialog. The required data source can be selected here by clicking on the corresponding entry. The selected entry then appears highlighted in colour.

All the channels of the chosen data source, that can be selected, are presented in the middle list.

The transferred channels are shown in the right hand list with their respective data sources. Activating 'OK' enables these channels to be processed further.

Channels in the middle list can be transferred to the right hand list via the control button  and . In order to do this one or more channels must first be selected in the middle list. To select multiple entries use the left mouse button together with the 'Shift' and 'Control' keys. To select all entries use the control button 'Select All'.

The sequence of the transferred entries in the right hand list can be changed via the control buttons  and . In this way, for example, the sequence of columns in a table is changed. To do this the required entry must first be clicked on.

An entry that appears in the right hand list by mistake can be removed using the control button .

With 'OK' the chosen channels with their selected data sources in the right hand list are accepted for further processing. The dialog is ended.

Cancel: The dialog is ended without transferring the selection. The procedure is halted.

## **2.6 Help Menu**

Commands for the Windows Online Help:

### **2.6.1 Additions to the manual**

The most recent changes and additions that have arisen after the printing of the manual.

### **2.6.2 Contents**

Starts the Online Help and shows the contents index of the Help topics.

### **2.6.3 Using Help**

Introduction to the Windows Help system.

### **2.6.4 Info**

Version number and Copyright of the current software version.

## 3 Appendix

### 3.1 Keyboard Operation

All Scroll and Zoom functions can also be easily operated using the keyboard. According to which unit is already active (the desired object can be also be selected with the keyboard by TAB), the following key functions are defined:

#### Graphic, Table and Report Windows

Ctrl+	Object Full Screen
Ctrl-	Reverse Object Full Screen
TAB	Change to Next Object

#### Graphic Window

+	Enlarge Graphic (Zoom)
-	Reduce Graphic (Un-zoom)
->	Scroll Right
<-	Scroll Left
Ctrl ->	Page Right
Ctrl <-	Page Left
End	Data End

#### Tables and Report Windows

Arrow Up	Scroll Up
Arrow Down	Scroll Down
Page Up	Large Step Back
Page Down	Large Step Forwards
Ctrl + Arrow Up	Scroll Back by 1 Day/Month/Year
Ctrl + Arrow Down	Scroll Forward by 1 Day/Month/Year
<-	Scroll Left
->	Scroll Right
Ctrl ->	Large Step to Right
Ctrl <-	Large Step to Left
Pos1	Go To Beginning of Table
End	Go To End of Table